

**Precision Power International, Inc.**

25108 Marguerite Parkway

Suite A-159

Mission Viejo, CA 92692 USA

T 949 951 6784

F 949 916 6733

sales@precisionpowerinternational.com**Specification Sheet** **LPS-2EVOLIS-BR**

Version 20100903

Certified - Instrumentation Grade: Instrument Power Protection System - IPPS (120 Vac) cUL / UL for two (2) each Bio-Rad Evolis™ Microplate Automation instrument combination with Controller

Catalog Number: **LPS-2EVOLIS-BR**

(1) Catalog Number LPS-2EVOLIS-BR supports the following Bio-Rad Evolis™ Microplate Automation (120 Vac) elements:

- | | |
|---------|---|
| Bio-Rad | (1) Workstation (Dell Computer and LCD Monitor or Laptop Computer)
(2) Evolis™ Microplate Automation instruments |
|---------|---|

PPI - Instrumentation Power Protection System (IPPS)**Specifications:** *(complete system protection)**(IPPS shown relative size)***(1) Bio-Rad Evolis™ Microplate Automation instrument shown with Computer & Monitor**

Instrument photo courtesy of Bio-Rad, Inc.

Design Installations: North America, Japan & Countries Utilizing NEMA Type Connectors (100-120 Vac 50/60 Hz)**Harmonization Code:** 8504.90.4000**Export License:** U.S. EAR (ECCN) EAR 99 - No License Required - NLR**Certification:** IEC/IEEE 62040 Laboratory and Instrumentation Grade
UL 1778, CSA 22.2-107, IEEE 62.41 CAT. A Lightning Safety**Performance:** Regenerated Power Technology (RPT), Digital Signal Processing (DSP)
True Sinewave Output, True Online, PWM Double Conversion (IGBT)
> 20,000 Hours MTBF