

**Precision Power International, Inc.**

25108 Marguerite Parkway

Suite A-159

Mission Viejo, CA 92692 USA

T 949 951 6784

F 949 916 6733

sales@precisionpowerinternational.com

Specification Sheet **LPS-CLUSSTA-IL**
Version 20090417

Certified - Instrumentation Grade:**Instrument Power Protection System (IPPS) for an Illumina Cluster Station (120 Vac) cUL****Catalog Number:** **LPS-CLUSSTA-IL (cUL)**

(1) Catalog Number LPS-CLUSSTA-IL supports the following Illumina (120 Vac) instrument elements:

| | |
|----------|--|
| Illumina | Cluster Station Workstation (Dell or HP computer and LCD Monitor) |
|----------|--|

PPI - Instrumentation Power Protection System (IPPS)**Specifications:** *Complete system protection**(IPPS shown relative size)***Illumina Cluster Station & Computer Workstation**

Instrument photo courtesy of Illumina, Inc.

Design Installations: North America, Japan & Countries Utilizing NEMA Type Connectors (100-120 Vac 50/60 Hz)**Harmonization Code:** 8504.90.4000**Export License:** U.S. EAR (ECCN) EAR 99 - No License Required - NLR**Certification:** IEC/IEEE 62040 Laboratory and Instrumentation Grade
UL 1778, CSA 22.2-107, IEEE 62.41 CAT. A Lightning Safety**Performance:** Regenerated Power Technology (RPT), Digital Signal Processing (DSP)
True Sinewave Output, True Online, PWM Double Conversion (IGBT)
> 20,000 Hours MTBF