

**Precision Power International, Inc.**

25108 Marguerite Parkway

Suite A-159

Mission Viejo, CA 92692 USA

T 949 951 6784

F 949 916 6733

sales@precisionpowerinternational.com**Specification Sheet LPS-SGARDBSC-BC**

Version 20100914

Certified - Instrumentation Grade: Laboratory / Instrument Power Protection System - IPPS (120 Vac) cUL for a Baker SterilGARD® III Advance° SG Series Class II, Type A2 (NSF/ANSI 49) Biological Safety Cabinet

Catalog Number: LPS-SGARDBSC-BC

(1) Catalog Number LPS-SGARDBSC supports one (1) of the following SterilGARD® III Advance° SG e3 (120 Vac) BSC units:

SG-303	(1) 36" nominal Class II, BSC, ETL Listed, or
SG-303a (HE)	(1) 36" nominal Class II, Type A2 (B-3) BSC, NSF/ANSI / UL, or
SG-403	(1) 48" nominal Class II, BSC, ETL Listed, or
SG-403a (HE)	(1) 48" nominal Class II, Type A2 (B-3) BSC, NSF/ANSI / UL, or
SG-503	(1) 60" nominal Class II, BSC, ETL Listed, or
SG-503a (HE)	(1) 60" nominal Class II, Type A2 (BSL-3) BSC, NSF/ANSI / UL, or
SG-603	(1) 72" nominal Class II, BSC, ETL Listed, or
SG-603a (HE)	(1) 72" nominal Class II, Type A2 (BSL-3) BSC, NSF/ANSI / UL

**PPI - Instrumentation Power Protection System (IPPS)**Specifications: **(complete system protection)***(IPPS shown relative size)***Baker SterilGARD® III Advance° (e3) SG Series Class II, Type A2 (NSF/ANSI 49) Biological Safety Cabinet**

instrument photo courtesy of The Baker Company

Design Installations: North America, Japan & Countries Utilizing NEMA Type Connectors (100-120 Vac 50/60 Hz)**Harmonization Code:** 8504.90.4000**Export License:** U.S. EAR (ECCN) EAR 99 - No License Required - NLR**Certification:** IEC/IEEE 62040 Laboratory and Instrumentation Grade
UL 1778, CSA 22.2-107, IEEE 62.41 CAT. A Lightning Safety**Performance:** Regenerated Power Technology (RPT), Digital Signal Processing (DSP)
True Sinewave Output, True Online, PWM Double Conversion (IGBT)
> 20,000 Hours MTBF